



Water Supply

Matrices with low concentrations of analytes for testing water supply, drinking water, or ground water. Standards are based on requirements of the United States Environmental Protection Agency Safe Drinking Water Act and may be used to satisfy PT requirements worldwide.



Water Supply PT Schedule

2026 Schedule

	Scheme #	Opens	Closes
Q	WS 354	Jan 12	Feb 26
	WS 355	Feb 9	Mar 26
	WS 356	Mar 9	Apr 23
Q	WS 357	Apr 13	May 28
	WS 358	May 11	Jun 25
	WS 359	Jun 8	Jul 23
Q	WS 360	Jul 13	Aug 27
	WS 361	Aug 10	Sep 24
	WS 362	Sep 8	Oct 23
Q	WS 363	Oct 9	Nov 23
	WS 364	Nov 2	Dec 17
	WS 365	Dec 7	Jan 21, 2027

2027 Schedule

	Scheme #	Opens	Closes
Q	WS 366	Jan 11	Feb 25
	WS 367	Feb 8	Mar 25
	WS 368	Mar 8	Apr 22
Q	WS 369	Apr 12	May 27
	WS 370	May 10	Jun 24
	WS 371	Jun 7	Jul 22
Q	WS 372	Jul 12	Aug 26
	WS 373	Aug 9	Sep 23
	WS 374	Sep 7	Oct 22
Q	WS 375	Oct 8	Nov 22
	WS 376	Nov 1	Dec 16
	WS 377	Dec 6	Jan 20, 2028

Schedule subject to change - see Waters ERA's website at eraqc.com

For the latest products and information, please visit us online at eraqc.com

Contents

CRM Certified Reference Material
PT Proficiency Testing
QR QuiK Response
RM Reference Material

All Waters ERA WS PTs open monthly (**M**), quarterly (**Q**), or biannually (**B**) unless otherwise noted.

Quarterly months are January, April, July, and October. Biannual months are January and July.

Description	CRM	PT	QR	Page
1,4-Dioxane	689	272 B	689QR	26
Ammonia as N	1359	1319 B	1359QR	25
Carbamates/Carbamoyloxime Pesticides	707	846 M	707QR	28
Chloral Hydrate	676	853 B	676QR	25
Chlordane	705	845 M	705QR	28
Chlorinated Acid Herbicides	704	851 M	704QR	30
Color	661C	859C Q	661CQR	26
Corrosivity	980	900 Q	980QR	26
Cyanide	983	556 M	983QR	25
Dioxin	663	857 Q	663QR	28
EDB/DBCP/TCP	706	847 M	706QR	28
Gasoline Additives	909	905 Q	909QR	26
Haloacetic Acids (HAA)	684	852 M	684QR	25
Halomethanes (THMs)	702	842 M	702QR	26
Hardness	693	555 M	693QR	24
Hexavalent Chromium	658	854 Q	658QR	24
Inorganic Disinfection #1	5272	5270 M	5272QR	25
Inorganic Disinfection #2	5262	5260 M	5262QR	25
Inorganics	698	591 M	698QR	24
Low-Level 1,2,3-TCP	682	596 B	682QR	28
Mercury	666	551 M	666QR	24
Metals	697	590 M	697QR	24

Description	CRM	PT	QR	Page
Nitrite	695	594 M	695QR	25
o-Phosphate Nutrients	667	558 M	667QR	25
Organic Carbon	669	557 M	669QR	25
PCBs as Decachlorobiphenyl	708	839 Q	708QR	28
Perchlorate	910	903 Q	910QR	25
Pesticides	709	850 M	709QR	28
PFAS Drinking Water	733	959 Q	733QR	27
pH	779	552 M	779QR	25
Regulated Volatiles	703	840 M	703QR	26
Residual Chlorine	696	593 M	696QR	25
Semivolatiles #1	690	848 M	690QR	28
Semivolatiles #2 Herbicides	691	849 M	691QR	30
Silica	785	902 Q	785QR	25
Solids Concentrate	5152	5150 M	5152QR	24
Surfactants-MBAS	784	901 Q	784QR	25
Toxaphene	700	844 M	700QR	28
Turbidity	699	592 M	699QR	26
Unregulated Volatiles	683	841 M	683QR	26
Uranium	930	858 Q	930QR	24
UV 254 Absorbance	662	904 Q	662QR	26
Vanadium	660	856 B	660QR	24

CRM: A reference material characterized by a metrologically valid procedure for one or more specified properties, accompanied by a reference material certificate that provides the value of the specified property, its associated uncertainty, and a statement of metrological traceability.

A complete listing of ERA's CRMs can be found on our Scope of Accreditation for general requirements for competence of reference material producers available at eraqc.com/Accreditations.

PT: A Proficiency Test (PT) is an analysis of what is often referred to as a blind sample or a sample with unknown concentrations of analytes for the purpose of evaluating a laboratory's analytical performance.

QR: Similar to a Proficiency Test, a QuiK Response (QR) is a sample with unknown concentrations. However, unlike a scheduled PT, QR is on-demand and available at any time. Plus, your results are returned within two business days. QuiK Response can be used as a bilateral PT as referenced in the IUPAC/CITAC guide: Selection and use of PT schemes for a limited number of participants – chemical analytical labs.

RM: A material, sufficiently homogeneous and stable with respect to one or more specified properties, which has been established to be fit for its intended use in a measurement process.

Minerals/Solids

Hardness

CRM Cat. #693	PT Cat. #555	M	QR Cat. #693QR
------------------	-----------------	---	-------------------

One 250 mL whole-volume bottle is ready to analyze.

Calcium	30-90 mg/L
Calcium hardness as CaCO ₃	75-225 mg/L
Total hardness as CaCO ₃	83-307 mg/L
Magnesium.....	2-20 mg/L
Sodium.....	12-50 mg/L

Inorganics

CRM Cat. #698	PT Cat. #591	M	QR Cat. #698QR
------------------	-----------------	---	-------------------

One 500 mL whole-volume bottle is ready to analyze. The CRM is also certified for sodium at 10-400 mg/L. For a sodium PT, order Hardness, Cat. #555.

Alkalinity as CaCO ₃	25-200 mg/L
Chloride.....	20-160 mg/L
Fluoride.....	1-8 mg/L
Nitrate as N.....	3-10 mg/L
Nitrate plus nitrite as N.....	3-10 mg/L
Potassium.....	10-40 mg/L
Specific conductance at 25 °C.....	130-1300 µmhos/cm
Sulfate.....	25-250 mg/L
Total dissolved solids (TDS) at 180 °C.....	100-1000 mg/L

Solids Concentrate

CRM Cat. #5152	PT Cat. #5150	M	QR Cat. #5152QR
-------------------	------------------	---	--------------------

One 24 mL screw-cap vial with a powder yields 1 liter after dilution.

Total filterable residue (TDS) at 180 °C.....	100-1000 mg/L
Total solids (TS) at 105 °C.....	123-1100 mg/L
Total suspended solids (TSS).....	23-100 mg/L



Trace Metals

Metals

CRM Cat. #697	PT Cat. #590	M	QR Cat. #697QR
------------------	-----------------	---	-------------------

One 15 mL screw-cap vial yields up to 2 liters after dilution. Use with ICP-OES, ICP-MS, and AA methods.

Aluminum.....	130-1000 µg/L
Antimony.....	6-50 µg/L
Arsenic.....	5-50 µg/L
Barium.....	500-3000 µg/L
Beryllium.....	2-20 µg/L
Boron.....	800-2000 µg/L
Cadmium.....	2-50 µg/L
Chromium.....	10-200 µg/L
Copper.....	50-2000 µg/L
Iron.....	100-1800 µg/L
Lead.....	5-100 µg/L
Manganese.....	40-900 µg/L
Molybdenum.....	15-130 µg/L
Nickel.....	10-500 µg/L
Selenium.....	10-100 µg/L
Silver.....	20-300 µg/L
Thallium.....	2-10 µg/L
Vanadium.....	50-1000 µg/L
Zinc.....	200-2000 µg/L

Mercury

CRM Cat. #666	PT Cat. #551	M	QR Cat. #666QR
------------------	-----------------	---	-------------------

One 15 mL screw-cap vial yields up to 1 liter after dilution. Use with CVAA, ICP-MS, or CVAFS methods.

Total mercury.....	0.5-10 µg/L
--------------------	-------------

Hexavalent Chromium

CRM Cat. #658	PT Cat. #854	Q	QR Cat. #658QR
------------------	-----------------	---	-------------------

One 15 mL screw-cap vial yields up to 2 liters after dilution.

Hexavalent chromium.....	5-50 µg/L
--------------------------	-----------

Uranium

CRM Cat. #930	PT Cat. #858	Q	QR Cat. #930QR
------------------	-----------------	---	-------------------

One 15 mL screw-cap vial yields up to 2 liters after dilution. Use with ICP-MS methods.

Uranium.....	3-104 µg/L
--------------	------------

Vanadium

CRM Cat. #660	PT Cat. #856	B	QR Cat. #660QR
------------------	-----------------	---	-------------------

One 15 mL screw-cap vial yields up to 2 liters after dilution. Designed to meet California ELAP requirements.

Vanadium.....	5-50 µg/L
---------------	-----------

B Waters ERA Vanadium PTs open in April and October.

Disinfection By-Products

Chloral Hydrate

CRM	PT	B	QR
Cat. #676	Cat. #853		Cat. #676QR

One 2 mL flame-sealed ampule yields in excess of 200 mL after dilution. Use with EPA Method 551, or other applicable method. Includes chloral hydrate at 4–30 µg/L.

B Waters ERA WS Chloral Hydrate PTs open in January and July.

Haloacetic Acids (HAA)

CRM	PT	M	QR
Cat. #684	Cat. #852		Cat. #684QR

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Method 552, or other applicable method. Includes all the analytes below at 5–50 µg/L.

Bromochloroacetic acid Dichloroacetic acid Monochloroacetic acid
Dibromoacetic acid Monobromoacetic acid Trichloroacetic acid

Inorganic Disinfection #1

CRM	PT	M	QR
Cat. #5272	Cat. #5270		Cat. #5272QR

One 24 mL screw-cap vial yields up to 4 liters after dilution.

Chlorate.....60–180 µg/L
Chlorite.....100–1000 µg/L

Inorganic Disinfection #2

CRM	PT	M	QR
Cat. #5262	Cat. #5260		Cat. #5262QR

One 24 mL screw-cap vial yields up to 4 liters after dilution.

Bromate.....7–50 µg/L
Bromide.....50–300 µg/L

Nutrients

Ammonia as N

CRM	PT	B	QR
Cat. #1359	Cat. #1319		Cat. #1359QR

One 15 mL screw-cap vial yields up to 1 liter after dilution.

Ammonia as N.....0.1–1 mg/L

B Waters ERA WS Ammonia as N PTs open in January and July.

Nitrite

CRM	PT	M	QR
Cat. #695	Cat. #594		Cat. #695QR

One 15 mL screw-cap vial yields up to 2 liters after dilution.

Nitrite as N.....0.4–2 mg/L

o-Phosphate Nutrients

CRM	PT	M	QR
Cat. #667	Cat. #558		Cat. #667QR

One 15 mL screw-cap vial yields up to 2 liters after dilution.

ortho-Phosphate as P.....0.5–5.5 mg/L

Miscellaneous Inorganic

Residual Chlorine

CRM	PT	M	QR
Cat. #696	Cat. #593		Cat. #696QR

One 2 mL flame-sealed ampule yields up to 2 liters after dilution.

Total residual chlorine.....0.5–3 mg/L
Free residual chlorine.....0.5–3 mg/L

Cyanide

CRM	PT	M	QR
Cat. #983	Cat. #556		Cat. #983QR

One 15 mL screw-cap vial yields up to 2 liters after dilution. Source material is free cyanide.

Free cyanide.....0.1–0.5 mg/L
Total cyanide.....0.1–0.5 mg/L
Cyanide.....0.1–0.5 mg/L

Organic Carbon

CRM	PT	M	QR
Cat. #669	Cat. #557		Cat. #669QR

One 15 mL screw-cap vial yields up to 2 liters after dilution.

Total organic carbon.....1.3–13 mg/L
Dissolved organic carbon.....1.3–13 mg/L

Perchlorate

CRM	PT	Q	QR
Cat. #910	Cat. #903		Cat. #910QR

One 15 mL screw-cap vial yields up to 1 liter after dilution.

Perchlorate.....4–20 µg/L

pH

CRM	PT	M	QR
Cat. #779	Cat. #552		Cat. #779QR

One 250 mL whole-volume bottle is ready to analyze.

pH.....5–10 units

Silica

CRM	PT	Q	QR
Cat. #785	Cat. #902		Cat. #785QR

One 60 mL poly bottle yields 1 liter after dilution.

Silica as SiO₂.....5–75 mg/L

Surfactants-MBAS

CRM	PT	Q	QR
Cat. #784	Cat. #901		Cat. #784QR

One 15 mL screw-cap vial yields up to 2 liters after dilution.

Surfactants-MBAS.....0.1–1 mg/L

Physical Property

Color			
CRM	PT	Q	QR
Cat. #661C	Cat. #859C		Cat. #661CQR

One 30 mL screw-cap bottle yields up to 200 mL after dilution.

Color10-75 PC units

Corrosivity			
CRM	PT	Q	QR
Cat. #980	Cat. #900		Cat. #980QR

One 500 mL whole-volume bottle is ready to analyze for corrosivity, calcium carbonate saturation, and Langelier Saturation Index.

Corrosivity-4 to +4 SI units

Turbidity			
CRM	PT	M	QR
Cat. #699	Cat. #592		Cat. #699QR

One 24 mL amber glass vial yields up to 1 liter after dilution. Use with nephelometric methods.

Turbidity0.5-8 NTU

UV 254 Absorbance			
CRM	PT	Q	QR
Cat. #662	Cat. #904		Cat. #662QR

One 15 mL screw-cap vial yields up to 1 liter after dilution.

UV 254 absorbance 0.05-0.7 cm-1

Volatile Organics

1,4-Dioxane			
CRM	PT	B	QR
Cat. #689	Cat. #272		Cat. #689QR

One 2 mL flame-sealed ampule yields 500 mL after dilution. Use with EPA method 522.

1,4-Dioxane0.1-10 µg/L

Gasoline Additives			
CRM	PT	Q	QR
Cat. #909	Cat. #905		Cat. #909QR

One 2 mL flame-sealed ampule yields in excess of 200 mL after dilution. Use with EPA Method 524.2, or other applicable method for gasoline additives/oxygenates. Contains all of the analytes below at 5-50 µg/L.

tert-Amyl methyl ether (TAME)	Ethyl tert-butyl ether (ETBE)	Trichlorofluoromethane (Freon® 11)
tert-Butyl alcohol	Methyl tert-butyl ether (MTBE)	Trichlorotrifluoroethane (Freon 113)
Di-isopropylether (DIPE)		

Halomethanes (THMs)			
CRM	PT	M	QR
Cat. #702	Cat. #842		Cat. #702QR

One 2 mL flame-sealed ampule yields in excess of 200 mL after dilution. Use with EPA Methods 502.2, 524.2, 551, or other applicable method. Contains all of the analytes below at 5-50 µg/L.

Bromodichloromethane	Chlorodibromomethane	Chloroform
Bromoform		

Regulated Volatiles			
CRM	PT	M	QR
Cat. #703	Cat. #840		Cat. #703QR

One 2 mL flame-sealed ampule yields in excess of 200 mL after dilution. Use with EPA Methods 502.2, 524.2, or other applicable method. Contains all of the analytes below at 2-50 µg/L.

Benzene	cis-1,2-Dichloroethylene	Toluene
Carbon tetrachloride	trans-1,2-Dichloroethylene	1,2,4-Trichlorobenzene
Chlorobenzene	1,2-Dichloropropane	1,1,1-Trichloroethane
1,2-Dichlorobenzene	Ethylbenzene	1,1,2-Trichloroethane
1,4-Dichlorobenzene	Methylene chloride	Trichloroethylene
1,2-Dichloroethane	Styrene	Vinyl chloride
1,1-Dichloroethylene	Tetrachloroethylene	Xylenes, total

Unregulated Volatiles			
CRM	PT	M	QR
Cat. #683	Cat. #841		Cat. #683QR

One 2 mL flame-sealed ampule yields in excess of 200 mL after dilution. Use with EPA Methods 502.2, 524.2, or other applicable method. Contains at least 60% of the analytes randomly selected from the list below at 2-50 µg/L.

Bromobenzene	1,3-Dichlorobenzene	4-Isopropyltoluene
Bromochloromethane	Dichlorodifluoromethane	Methyl tert-butyl ether (MTBE)
Bromomethane	1,1-Dichloroethane	Naphthalene
n-Butylbenzene	1,3-Dichloropropane	n-Propylbenzene
sec-Butylbenzene	2,2-Dichloropropane	1,1,1,2-Tetrachloroethane
tert-Butylbenzene	1,1-Dichloropropene	1,1,2,2-Tetrachloroethane
Chloroethane	cis-1,3-Dichloropropene	1,2,3-Trichlorobenzene
Chloromethane	trans-1,3-Dichloropropene	1,2,3-Trichloropropane
2-Chlorotoluene	Fluorotrichloromethane	1,2,4-Trimethylbenzene
4-Chlorotoluene	Hexachlorobutadiene	1,3,5-Trimethylbenzene
Dibromomethane	Isopropylbenzene	

Per- and Polyfluoroalkyl Substances (PFAS)

PFAS in Drinking Water

CRM Cat. #733	PT Cat. #959	Q	QR Cat. #733QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields in excess of 1.5 L after dilution. The sample is designed for LC/MS/MS methods for analyzing potable water, specifically EPA Methods 533, 537 and 537.1. The diluted standard is certified for the 32 analytes listed below.

Perfluorobutanoic acid, PFBA.....	20-200 ng/L
Perfluoropentanoic acid, PFPeA.....	20-200 ng/L
Perfluorohexanoic acid, PFHxA.....	20-200 ng/L
Perfluoroheptanoic acid, PFHpA.....	20-200 ng/L
Perfluorooctanoic acid, PFOA.....	20-200 ng/L
Perfluorononanoic acid, PFNA.....	20-200 ng/L
Perfluorodecanoic acid, PFDA.....	20-200 ng/L
Perfluoroundecanoic acid, PFUDA.....	20-200 ng/L
Perfluorododecanoic acid, PFDoA.....	20-200 ng/L
Perfluorotridecanoic acid, PFTrDA.....	20-200 ng/L
Perfluorotetradecanoic acid, PFTeDA.....	20-200 ng/L
Perfluorobutanesulfonic acid, PFBS.....	20-200 ng/L
Perfluoropentanesulfonic acid, PFPeS.....	20-200 ng/L
Perfluorohexanesulfonic acid, PFHxS.....	20-200 ng/L
Perfluoroheptanesulfonic acid, PFHpS.....	20-200 ng/L
Perfluorooctanesulfonic acid, PFOS.....	20-200 ng/L
Perfluorononanesulfonic acid, PFNS.....	20-200 ng/L
Perfluorodecanesulfonic acid, PFDS.....	20-200 ng/L
4:2 fluorotelomersulfonic acid, 4:2 FTS.....	20-200 ng/L
6:2 fluorotelomersulfonic acid, 6:2 FTS.....	20-200 ng/L
8:2 fluorotelomersulfonic acid, 8:2 FTS.....	20-200 ng/L
Perfluorooctanesulfonamide, PFOSA.....	20-200 ng/L
N-ethyl perfluorooctanesulfonamidoacetic acid, NETFOSAA.....	20-200 ng/L
N-methyl perfluorooctanesulfonamidoacetic acid, NMEFOSAA.....	20-200 ng/L
Hexafluoropropylene oxide dimer acid, HFPO-DA.....	20-200 ng/L
4,8-dioxa-3H-perfluorononanoic acid, ADONA.....	20-200 ng/L
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid, 9Cl-PF3ONS.....	20-200 ng/L
11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid, 11Cl-PF3OUdS.....	20-200 ng/L
Perfluoro-4-methoxybutanoic acid, PFMBA.....	20-200 ng/L
Perfluoro-3-methoxypropanoic acid, PFMPA.....	20-200 ng/L
Perfluoro(2-ethoxyethane) sulfonic acid, PFEESA.....	20-200 ng/L
Nonafluoro-3,6-dioxaheptanoic acid, NFDHA.....	20-200 ng/L

PFAS Secondary Source Standard

Standard is suitable for various applications, including Internal Calibration Verification (ICV), Laboratory Control Sample (LCS), Matrix Spike (MS), and Limit of Quantitation (LOQ) studies.

NEW!

Drinking Water

CRM

Cat. #PFAS10002

One 2 mL flame-sealed ampule with 1.5 mL of PFAS standard containing 29 analytes at 50 ng/mL. The standard is suitable for matrices to include, but not limited to, drinking water and compatible with methods EPA 533, EPA 537, EPA 537.1 and other comparable methods.

Perfluorobutanoic acid, PFBA.....	50 ng/mL
Perfluoropentanoic acid, PFPeA.....	50 ng/mL
Perfluorohexanoic acid, PFHxA.....	50 ng/mL
Perfluoroheptanoic acid, PFHpA.....	50 ng/mL
Perfluorooctanoic acid, PFOA.....	50 ng/mL
Perfluorononanoic acid, PFNA.....	50 ng/mL
Perfluorodecanoic acid, PFDA.....	50 ng/mL
Perfluoroundecanoic acid, PFUDA.....	50 ng/mL
Perfluorododecanoic acid, PFDoA.....	50 ng/mL
Perfluorotridecanoic acid, PFTrDA.....	50 ng/mL
Perfluorotetradecanoic acid, PFTeDA.....	50 ng/mL
Perfluorobutanesulfonic acid, PFBS.....	50 ng/mL
Perfluoropentanesulfonic acid, PFPeS.....	50 ng/mL
Perfluorohexanesulfonic acid, PFHxS.....	50 ng/mL
Perfluoroheptanesulfonic acid, PFHpS.....	50 ng/mL
Perfluorooctanesulfonic acid, PFOS.....	50 ng/mL
4:2 fluorotelomersulfonic acid, 4:2FTS.....	50 ng/mL
6:2 fluorotelomersulfonic acid, 6:2FTS.....	50 ng/mL
8:2 fluorotelomersulfonic acid, 8:2FTS.....	50 ng/mL
N-ethyl perfluorooctanesulfonamidoacetic acid, NETFOSAA.....	50 ng/mL
N-methyl perfluorooctanesulfonamidoacetic acid, NMEFOSAA.....	50 ng/mL
Hexafluoropropylene oxide dimer acid, HFPO-DA.....	50 ng/mL
4,8-dioxa-3H-perfluorononanoic acid, ADONA.....	50 ng/mL
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid, 9Cl-PF3ONS.....	50 ng/mL
11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid, 11Cl-PF3OUdS.....	50 ng/mL
Perfluoro-4-methoxybutanoic acid, PFMBA.....	50 ng/mL
Perfluoro-3-methoxypropanoic acid, PFMPA.....	50 ng/mL
Perfluoro(2-ethoxyethane) sulfonic acid, PFEESA.....	50 ng/mL
Nonafluoro-3,6-dioxaheptanoic acid, NFDHA.....	50 ng/mL



Learn more about WS products

Pesticides

Pesticides

CRM Cat. #709	PT Cat. #850	M	QR Cat. #709QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 505, 507, 508, 525, or other applicable method for organochlorine, nitrogen, and organophosphorus pesticides. Each standard contains at least 14 analytes randomly selected from the list below at 0.2-20 µg/L.

Alachlor	Heptachlor	Metribuzin
Aldrin	Heptachlor epoxide (beta)	Molinate (ordram)
Atrazine	Hexachlorobenzene	Prometon
Bromacil	Hexachlorocyclopentadiene	Propachlor
Butachlor	Lindane (gamma-BHC)	Simazine
Diazinon	Methoxychlor	Thiobencarb
Dieldrin	Metolachlor	Trifluralin
Endrin		

Carbamate/Carbamoxymoxime Pesticides

CRM Cat. #707	PT Cat. #846	M	QR Cat. #707QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 531.1, 531.2, 632, or other applicable method. Each standard contains at least 8 of the analytes below at 15-150 µg/L.

Aldicarb	Carbaryl	Methiocarb
Aldicarb sulfone	Carbofuran	Methomyl
Aldicarb sulfoxide	3-Hydroxycarbofuran	Oxamyl
Baygon		

EDB/DBCP/TCP

CRM Cat. #706	PT Cat. #847	M	QR Cat. #706QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields in excess of 200 mL after dilution. Use with EPA Methods 504, 551, or other applicable method. Each lot contains all analytes below at 0.05-2 µg/L.

1,2-Dibromo-3-chloropropane (DBCP)
Ethylene dibromide (EDB)
1,2,3-Trichloropropane (1,2,3-TCP)

Low-Level 1,2,3-TCP

CRM Cat. #682	PT Cat. #596	B	QR Cat. #682QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields 100 mL after dilution. Use with California method SRL 524M, or other applicable method. Each standard contains 1,2,3-Trichloropropane (TCP) at 5-100 ng/L after dilution.

B Low-Level 1,2,3-TCP available in January and July.

Chlordane

CRM Cat. #705	PT Cat. #845	M	QR Cat. #705QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 505, 508, 525, or other applicable method. Each standard contains technical chlordane at 2-20 µg/L.

Toxaphene

CRM Cat. #700	PT Cat. #844	M	QR Cat. #700QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 505, 508, 525, or other applicable method. Each standard contains toxaphene at 2-20 µg/L.

Semivolatile Organics

Dioxin

CRM Cat. #663	PT Cat. #857	Q	QR Cat. #663QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 613, 1613, 8280, 8290, or other applicable method. Each standard contains 2,3,7,8-TCDD at 20-100 µg/L.

PCBs as Decachlorobiphenyl

CRM Cat. #708	PT Cat. #839	Q	QR Cat. #708QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Quantitative Method 508A. This standard can also be used for aroclor identification and quantification using EPA Methods 505, 508, 508.1, or other applicable method. Includes an aroclor randomly selected from the list below at 0.5-5 µg/L as decachlorobiphenyl.

Aroclor 1016	Aroclor 1242	Aroclor 1254
Aroclor 1221	Aroclor 1248	Aroclor 1260
Aroclor 1232		

Semivolatiles #1

CRM Cat. #690	PT Cat. #848	M	QR Cat. #690QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 506, 525, 550, or other applicable method for PAHs, phthalates, and adipates. Each standard contains benzo(a)pyrene, bis(2-ethylhexyl)adipate, and bis(2-ethylhexyl)phthalate plus at least 13 additional analytes, selected from the list below, at 0.2-50 µg/L.

Acenaphthene	Butyl benzyl phthalate	bis(2-Ethylhexyl)phthalate
Acenaphthylene	Chrysene	Fluoranthene
Anthracene	Dibenz(a,h)anthracene	Fluorene
Benzo(a)anthracene	Di-n-butyl phthalate	Indeno(1,2,3-cd)pyrene
Benzo(b)fluoranthene	Diethyl phthalate	Naphthalene
Benzo(k)fluoranthene	Dimethyl phthalate	Phenanthrene
Benzo(g,h,i)perylene	Di-n-octyl phthalate	Pyrene
Benzo(a)pyrene	bis(2-Ethylhexyl)adipate	

Naphthalene is not within the EPA/NELAC range. Use the Unregulated Volatiles standard (page 27 for this compound in the EPA/NELAC range.

Tackle Your Most Stringent PFAS Limits and Get Ready to Conquer Your Analytical Challenges With the Waters LC-MS/MS Workflow Solutions

From sample prep to outcome-based professional services training and [proficiency testing](#), Waters is committed to revolutionizing your PFAS analysis. Partner with Waters PFAS experts to strengthen your analytical game plan and achieve detection limits as precise as single-digit ppq with the Xevo™ TQ Absolute. Safeguard your analysis against contamination and control interference with our PFAS LC kits and Oasis™ WAX SPE.

Sample Preparation	Data Management	LC-MS/MS	Complete Solutions
	 	 	 

Browse our PFAS solutions for comprehensive workflows for sample extraction, screening, quantification, and reporting of PFAS in water, environmental, food, or other complex matrices with developed methods to help you stay ahead of challenging PFAS analyses.



LEARN MORE AT
waters.com/PFAS

Herbicides

Chlorinated Acid Herbicides

CRM Cat. #704	PT Cat. #851	M	QR Cat. #704QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 515.1, 515.2, 515.3, 515.4, 555, or other applicable method. All lots include at least 10 analytes from the list below at 1–120 µg/L.

Acifluorfen	Dalapon	4-Nitrophenol
Bentazon	Dicamba	Pentachlorophenol
Chloramben	3,5-Dichlorobenzoic acid	Picloram
2,4-D	Dichlorprop	2,4,5-T
2,4-DB	Dinoseb	2,4,5-TP (silvex)
Dacthal diacid (DCPA)		

Semivolatiles #2 Herbicides

CRM Cat. #691	PT Cat. #849	M	QR Cat. #691QR
------------------	-----------------	---	-------------------

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with EPA Methods 547, 548, 549, or other applicable method. Each standard contains all the analytes below at 8–800 µg/L.

Diquat	Glyphosate	Paraquat
Endothall		

Evolution in Dioxin Analysis Technology



Waters provides industry leading technology to address complex environmental challenges affecting human health.

Dioxin analysis is particularly demanding due to low level regulatory exposure limits and complex sample matrices. With the adoption of atmospheric pressure chemical ionization with tandem mass spectrometry (APCI-MS/MS) as an acceptable alternative, your operational efficiency and analytical quality will improve as you experience:

- Improved robustness and sensitivity for increased productivity
- Improvements in sample preparation efficiency
- Versatile system capable of enhanced operation for SVOC and other analyses
- Widely compatible with carrier gas options including nitrogen



LEARN MORE AT [waters.com/dioxins](https://www.waters.com/dioxins)